



MALAYSIAN STANDARD

MS CISPR 11:2012

**Industrial, scientific and medical equipment -
Radio-frequency disturbance characteristics -
Limits and methods of measurement
(Second revision)
(CISPR 11:2010, IDT)**

ICS: 33.100.10

Descriptors: industrial, scientific, medical radio-frequency equipment, electromagnetic disturbance characteristics, limit, method of measurement

© Copyright 2012

DEPARTMENT OF STANDARDS MALAYSIA

DEVELOPMENT OF MALAYSIAN STANDARDS

The **Department of Standards Malaysia (STANDARDS MALAYSIA)** is the national standards and accreditation body of Malaysia.

The main function of STANDARDS MALAYSIA is to foster and promote standards, standardisation and accreditation as a means of advancing the national economy, promoting industrial efficiency and development, benefiting the health and safety of the public, protecting the consumers, facilitating domestic and international trade and furthering international cooperation in relation to standards and standardisation.

Malaysian Standards (MS) are developed through consensus by committees which comprise balanced representation of producers, users, consumers and others with relevant interests, as may be appropriate to the subject at hand. To the greatest extent possible, Malaysian Standards are aligned to or are adoption of international standards. Approval of a standard as a Malaysian Standard is governed by the Standards of Malaysia Act 1996 [Act 549]. Malaysian Standards are reviewed periodically. The use of Malaysian Standards is voluntary except in so far as they are made mandatory by regulatory authorities by means of regulations, local by-laws or any other similar ways.

STANDARDS MALAYSIA has appointed **SIRIM Berhad** as the agent to develop, distribute and sell the Malaysian Standards.

For further information on Malaysian Standards, please contact:

Department of Standards Malaysia
Ministry of Science, Technology and Innovation
Level 1 & 2, Block 2300, Century Square
Jalan Usahawan
63000 Cyberjaya
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 8318 0002
Fax: 60 3 8319 3131
<http://www.standardsmalaysia.gov.my>

E-mail: central@standardsmalaysia.gov.my

OR **SIRIM Berhad**
(Company No. 367474 - V)
1, Persiaran Dato' Menteri
Section 2, P.O. Box 7035
40700 Shah Alam
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 5544 6000
Fax: 60 3 5510 8095
<http://www.sirim.my>

E-mail: msonline@sirim.my

CONTENTS

	Page
Committee representation.....	iv
National foreword.....	v
Foreword.....	vi
Introduction.....	ix
1 Scope.....	9
2 Normative references.....	9
3 Terms and definitions	10
4 Frequencies designated for ISM use	11
5 Classification of ISM equipment	12
5.1 Information for the user	12
5.2 Separation into groups.....	12
5.3 Division into classes	12
6 Limits of electromagnetic disturbances.....	13
6.1 General	13
6.2 Group 1 equipment measured on a test site	13
6.2.1 Limits of terminal disturbance voltage	13
6.2.2 Limits of electromagnetic radiation disturbance	14
6.3 Group 2 equipment measured on a test site	17
6.3.1 Limits of terminal disturbance voltage	17
6.3.2 Limits of electromagnetic radiation disturbance	19
6.4 Group 1 and group 2 class A equipment measured <i>in situ</i>	27
6.4.1 Limits of terminal disturbance voltage	27
6.4.2 Limits of electromagnetic radiation disturbance	27
7 Measurement requirements.....	29
7.1 General	29
7.2 Ambient noise.....	29
7.3 Measuring equipment	29
7.3.1 Measuring instruments.....	29
7.3.2 Artificial mains network	30
7.3.3 Voltage probe	30
7.3.4 Antennas	30
7.3.5 Artificial hand.....	31
7.4 Frequency measurement	31
7.5 Configuration of equipment under test.....	31
7.5.1 General	31
7.5.2 Interconnecting cables	31
7.5.3 Connection to the electricity supply network on a test site.....	32
7.6 Load conditions of equipment under test.....	33
7.6.1 General	33
7.6.2 Medical equipment.....	33
7.6.3 Industrial equipment	34
7.6.4 Scientific, laboratory and measuring equipment.....	35
7.6.5 Microwave cooking appliances	35

CONTENTS (*continued*)

	Page
7.6.6 Other equipment in the frequency range 1 GHz to 18 GHz.....	35
7.6.7 Single and multiple-zone induction cooking appliances.....	36
7.6.8 Electric welding equipment.....	36
7.7 Recording of test-site measurement results	37
7.7.1 General	37
7.7.2 Conducted emissions.....	37
7.7.3 Radiated emissions.....	37
8 Special provisions for test site measurements (9 kHz to 1 GHz)	37
8.1 Ground planes.....	37
8.2 Measurement of mains terminal disturbance voltage	37
8.2.1 General	37
8.2.2 Handheld equipment which are normally operated without an earth connection.....	38
8.3 Radiation test site for 9 kHz to 1 GHz.....	38
8.3.1 General	38
8.3.2 Validation of the radiation test site (9 kHz to 1 GHz).....	39
8.3.3 Disposition of equipment under test (9 kHz to 1 GHz).....	39
8.3.4 Radiation measurements (9 kHz to 1 GHz).....	39
8.4 Alternative radiation test sites for the frequency range 30 MHz to 1 GHz	39
9 Radiation measurements: 1 GHz to 18 GHz	40
9.1 Test arrangement	40
9.2 Receiving antenna.....	40
9.3 Validation and calibration of test site.....	40
9.4 Measuring procedure.....	40
10 Measurement <i>in situ</i>	40
11 Safety precautions	41
12 Assessment of conformity of equipment.....	41
12.1 General.....	41
12.2 Statistical assessment of compliance of series produced equipment.....	41
12.3 Equipment in small-scale production	42
12.4 Equipment produced on an individual basis.....	42
12.5 Measurement uncertainty.....	42
13 Figures and flowcharts.....	42
Annex A (informative) Examples of equipment classification.....	46
Annex B (informative) Precautions to be taken in the use of a spectrum analyzer (see 7.3.1)	48
Annex C (normative) Measurement of electromagnetic radiation disturbance in the presence of signals from radio transmitters	49
Annex D (informative) Propagation of interference from industrial radio-frequency equipment at frequencies between 30 MHz and 300 MHz.....	50
Annex E (informative) Recommendations of CISPR for protection of certain radio services in particular areas.....	51
Annex F (informative) Frequency bands allocated for safety-related radio services	52
Annex G (informative) Frequency bands allocated for sensitive radio services	53
Bibliography.....	54

CONTENTS (*concluded*)

	Page
Figure 1 – Test site	42
Figure 2 – Minimum size of metal ground plane	43
Figure 3 – Disposition of medical (capacitive type) and dummy load (see 7.6.2.1)	43
Figure 4 – Circuit for disturbance voltage measurements on mains supply (see 7.3.3).....	44
Figure 5 – Decision tree for the measurement of emissions from 1 GHz to 18 GHz of class B, group 2 ISM equipment operating at frequencies above 400 MHz	44
Figure 6 – Artificial hand, RC element (see 7.3.5).....	45
Table 1 – Frequencies in the radio-frequency (RF) range designated by ITU for use as fundamental ISM frequencies	12
Table 2 – Mains terminal disturbance voltage limits for class A group 1 equipment measured on a test site	14
Table 3 – Mains terminal disturbance voltage limits for class B group 1 equipment measured on a test site	14
Table 4 – Electromagnetic radiation disturbance limits for class A group 1 equipment measured on a test site	15
Table 5 – Electromagnetic radiation disturbance limits for class B group 1 equipment measured on a test site	16
Table 6 – Mains terminal disturbance voltage limits for class A group 2 equipment measured on a test site	18
Table 7 – Mains terminal disturbance voltage limits for class B group 2 equipment measured on a test site	18
Table 8 – Mains terminal disturbance voltage limits for induction cooking appliances	19
Table 9 – Electromagnetic radiation disturbance limits for class A group 2 equipment measured on a test site	21
Table 10 – Electromagnetic radiation disturbance limits for class A EDM and arc welding equipment measured on a test site	22
Table 11 – Electromagnetic radiation disturbance limits for class B group 2 equipment measured on a test site	23
Table 12 – Limits of the magnetic field strength for induction cooking appliances intended for commercial use	24
Table 13 – Limits of the magnetic field induced current in a 2 m loop antenna for induction cooking appliances for domestic use	25
Table 14 – Electromagnetic radiation disturbance peak limits for group 2 equipment producing CW type disturbances and operating at frequencies above 400 MHz.....	26
Table 15 – Electromagnetic radiation disturbance peak limits for class B group 2 equipment producing fluctuating disturbances other than CW and operating at frequencies above 400 MHz	26
Table 16 – Electromagnetic radiation disturbance weighted limits for class B group 2 equipment producing fluctuating disturbances other than CW and operating at frequencies above 400 MHz	27
Table 17 – Electromagnetic radiation disturbance limits for class A group 1 equipment measured <i>in situ</i>	27
Table 18 – Electromagnetic radiation disturbance limits for class A group 2 equipment measured <i>in situ</i>	28
Table 19 – The non-central <i>t</i> -distribution factor <i>k</i> as a function of the sample size <i>n</i>	41
Table E.1 – Limits for electromagnetic radiation disturbances for <i>in situ</i> measurements to protect specific safety-related radio services in particular areas	51

Committee representation

The Industry Standards Committee on Generation, Transmission and Distribution of Energy (ISC E) under whose authority this Malaysian Standard was adopted, comprises representatives from the following organisations:

Association of Consulting Engineers Malaysia
Atomic Energy Licensing Board
Department of Standards Malaysia
Federation of Malaysian Manufacturers
Jabatan Kerja Raya Malaysia
Malaysian Association of Standards Users
Malaysian Cable Manufacturers Association
Malaysian Electrical Appliances and Distributors Association
Malaysian Green Technology Corporation
Ministry of Domestic Trade, Co-operatives and Consumerism
Ministry of International Trade and Industry
Persatuan Kontraktor Elektrikal dan Mekanikal Melayu Malaysia
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Komunikasi dan Multimedia Malaysia
Suruhanjaya Tenaga
Tenaga Nasional Berhad
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
Universiti Teknologi Malaysia

The Technical Committee on Electromagnetic Compatibility which supervised the adoption of the IEC Standard as Malaysian Standard consists of representatives from the following organisations:

Federation of Malaysian Manufacturers
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Komunikasi dan Multimedia Malaysia
Suruhanjaya Tenaga
Tenaga Nasional Berhad
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
TNB Research Sdn Bhd
Universiti Teknologi Malaysia

The Working Group on EMC-CISPR-IEC which recommended the adoption of the IEC Standard as Malaysian Standard consists of representatives from the following organisations:

Institut Penyelidikan Sains dan Teknologi Pertahanan
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Tenaga
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
TNB Research Sdn Bhd

NATIONAL FOREWORD

The adoption of the IEC Standard as a Malaysian Standard was recommended by the Working Group on EMC-CISPR-IEC under the authority of the Industry Standards Committee on Generation, Transmission and Distribution of Energy.

This Malaysian Standard is the second revision of MS CISPR 11, *Industrial, scientific and medical (ISM) radio-frequency equipment - Electromagnetic disturbance characteristics - Limits and methods of measurement*.

This Malaysian Standard is identical with CISPR 11:2010, *Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement*, published by the International Electrotechnical Commission (IEC). However, for the purposes of this Malaysian Standard, the following apply:

- a) in the source text, "this International Standard" should read "this Malaysian Standard";
and
- b) the comma which is used as a decimal sign (if any), to read as a point.

This Malaysian Standard cancels and replaces MS CISPR 11:2005.

Compliance with a Malaysian Standard does not of itself confer immunity from legal obligations.

NOTE. IDT on the front cover indicates an identical standard i.e. a standard where the technical content, structure, and wording (or is an identical translation) of a Malaysian Standard is exactly the same as in an International Standard or is identical in technical content and structure although it may contain the minimal editorial changes specified in clause 4.2 of ISO/IEC Guide 21-1.