



MALAYSIAN STANDARD

MS ISO 11138-3:2010

STERILIZATION OF HEALTH CARE PRODUCTS - BIOLOGICAL INDICATORS - PART 3: BIOLOGICAL INDICATORS FOR MOIST HEAT STERILIZATION PROCESSES (ISO 11138-3:2006, IDT)

ICS: 11.080.01

Descriptors: medical device, sterilization, biological indicators, moist heat, requirements

© Copyright 2010

DEPARTMENT OF STANDARDS MALAYSIA

DEVELOPMENT OF MALAYSIAN STANDARDS

The **Department of Standards Malaysia (STANDARDS MALAYSIA)** is the national standards and accreditation body of Malaysia.

The main function of STANDARDS MALAYSIA is to foster and promote standards, standardisation and accreditation as a means of advancing the national economy, promoting industrial efficiency and development, benefiting the health and safety of the public, protecting the consumers, facilitating domestic and international trade and furthering international cooperation in relation to standards and standardisation.

Malaysian Standards (MS) are developed through consensus by committees which comprise balanced representation of producers, users, consumers and others with relevant interests, as may be appropriate to the subject at hand. To the greatest extent possible, Malaysian Standards are aligned to or are adoption of international standards. Approval of a standard as a Malaysian Standard is governed by the Standards of Malaysia Act 1996 [Act 549]. Malaysian Standards are reviewed periodically. The use of Malaysian Standards is voluntary except in so far as they are made mandatory by regulatory authorities by means of regulations, local by-laws or any other similar ways.

STANDARDS MALAYSIA has appointed **SIRIM Berhad** as the agent to develop, distribute and sell the Malaysian Standards.

For further information on Malaysian Standards, please contact:

Department of Standards Malaysia

Ministry of Science, Technology and Innovation
Level 1 & 2, Block 2300, Century Square
Jalan Usahawan
63000 Cyberjaya
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 8318 0002

Fax: 60 3 8319 3131

<http://www.standardsmalaysia.gov.my>

E-mail: central@standardsmalaysia.gov.my

OR

SIRIM Berhad

(Company No. 367474 - V)
1, Persiaran Dato' Menteri
Section 2
40000 Shah Alam
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 5544 6000

Fax: 60 3 5510 8095

<http://www.sirim.my>

E-mail: msonline@sirim.my

CONTENTS

	Page
Committee representation	ii
National foreword	iii
Foreword.....	iv
Introduction	v
1 Scope.....	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements	1
5 Test organism.....	2
6 Suspension	2
7 Carrier and primary packaging.....	2
8 Inoculated carriers and biological indicators	2
9 Population and resistance.....	2
Table B.1 Examples of data collected for regression analysis	5
Table B.2 Examples of calculations for slope	6
Annex A Method for determination of resistance to moist heat sterilization.....	4
Annex B Calculation of z value and correlation coefficient, r^2	5
Bibliography	8

Committee representation

The Industry Standards Committee on Medical Devices and Facilities for Healthcare (ISC R) under whose authority this Malaysian Standard was adopted, comprises representatives from the following organisations:

Association of Malaysia Medical Industries
Atomic Energy Licensing Board
Biomedical Engineering Association Malaysia
Department of Standards Malaysia
Federation of Malaysian Manufacturers
Institute for Medical Research
Malaysia Medical Device Association
Malaysian Association of Standards Users
Malaysian Medical Association
Malaysian Nuclear Agency
Malaysian Organisation of Pharmaceutical Industries
Malaysian Rubber Board
Malaysian Rubber Export Promotion Council
Ministry of Health Malaysia
Ministry of Health Malaysia (Medical Device Bureau)
Pharmaceutical Association of Malaysia
Radiation Physics, Biophysics and Medical Physics Sub-Group of Institute of Physics Malaysia
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Universiti Kebangsaan Malaysia
Universiti Teknologi Malaysia

The Technical Committee on Quality Management and Corresponding General Aspects for Medical Devices which recommended the adoption of the ISO Standard as Malaysian Standard consists of representatives from the following organisations:

Association of Private Hospitals of Malaysia
Biomedical Engineering Association Malaysia
Federation of Malaysian Manufacturers
Institute of Quality Malaysia
Malaysia Medical Device Association
Malaysian Medical Association
Malaysian Nuclear Agency
Malaysian Rubber Export Promotion Council
Malaysian Rubber Glove Manufacturers' Association
Ministry of Health Malaysia
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
SterilGamma (M) Sdn Bhd

NATIONAL FOREWORD

The adoption of the ISO Standard as a Malaysian Standard was recommended by the Technical Committee on Quality Management and Corresponding General Aspects for Medical Devices under the authority of the Industry Standards Committee on Medical Devices and Facilities for Healthcare.

This Malaysian Standard is identical with ISO 11138-3:2006, *Sterilization of health care products - Biological indicators - Part 3: Biological indicators for moist heat sterilization processes*, published by the International Organization for Standardization (ISO). However, for the purposes of this Malaysian Standard, the following apply:

- a) in the source text, "this International Standard" should read "this Malaysian Standard";
- b) the comma which is used as a decimal sign (if any), to read as a point; and
- c) reference to International Standards should be replaced by corresponding Malaysian Standards as follows:

Referenced International Standards

ISO 11138-1:2006, *Sterilization of health care products - Biological indicators - Part 1: General requirements*

Corresponding Malaysian Standards

MS ISO 11138-1:2011, *Sterilization of health care products - Biological indicators - Part 1: General requirements*

Compliance with a Malaysian Standard does not of itself confer immunity from legal obligations.

NOTE. IDT on the front cover indicates an identical standard i.e. a standard where the technical content, structure, and wording (or is an identical translation) of a Malaysian Standard is exactly the same as in an International Standard or is identical in technical content and structure although it may contain the minimal editorial changes specified in clause 4.2 of ISO/IEC Guide 21-1.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11138-3 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 11138-3:1995), which has been technically revised.

ISO 11138 consists of the following parts, under the general title *Sterilization of health care products — Biological indicators*:

- *Part 1: General requirements*
- *Part 2: Biological indicators for ethylene oxide sterilization processes*
- *Part 3: Biological indicators for moist heat sterilization processes*
- *Part 4: Biological indicators for dry heat sterilization processes*
- *Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes*

Introduction

ISO 11138-1 specifies production, labelling, test methods and performance requirements for the manufacture of biological indicators including inoculated carriers and suspensions intended for use in validation and monitoring of sterilization processes. This part of ISO 11138 gives specific requirements for those biological indicators intended for use in moist heat sterilization processes.

The intent of providing requirements in the ISO 11138 series of International Standards is to provide general requirements and requirements for test methods. This series of International Standards represents the current “state-of-the-art” according to the experts representing manufacturers, users and regulatory authorities involved in developing the standard. The intent is not to promote the use of biological indicators where such use is not advised, but to provide common requirements for the production of those biological indicators known to be in use today.

Standards exist providing requirements for the validation and control of moist heat sterilization (see ISO 17665).

NOTE Some countries or regions may have published standards covering requirements for sterilization or biological indicators.

Advice on selection, use and interpretation of results when using biological indicators can be found in ISO 14161.

Sterilization of health care products — Biological indicators —

Part 3: Biological indicators for moist heat sterilization processes

1 Scope

This part of ISO 11138 provides specific requirements for test organisms, suspensions, inoculated carriers, biological indicators and test methods intended for use in assessing the performance of sterilization processes employing moist heat as the sterilizing agent.

Moist heat as the sterilizing agent is defined in this part of ISO 11138 as dry saturated steam. While air-steam mixtures may be used in moist heat sterilization processes, the methods and performance requirements of this part of ISO 11138 might not be applicable for biological indicators used in such processes.

NOTE 1 Requirements for validation and control of moist heat sterilization processes are provided by ISO 17665.

NOTE 2 National or regional regulations may provide requirements for work place safety.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11138-1:2006, *Sterilization of health care products — Biological indicators — Part 1: General requirements*

ISO 18472, *Sterilization of health care products — Biological and chemical indicators — Test equipment*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11138-1 apply.

4 General requirements

The requirements of ISO 11138-1 apply.

5 Test organism

5.1 The test organisms shall be spores of *Geobacillus stearothermophilus* or other strains of microorganism of demonstrated equivalent performance as required by this part of ISO 11138.

NOTE 1 *Bacillus stearothermophilus* has been reclassified as *Geobacillus stearothermophilus*.

NOTE 2 *Geobacillus stearothermophilus* ATCC 7953 (NCTC 10007, DSM 22 and CIP 52.81), ATCC 12980 (equivalent to NRRL B-4419), have been found to be suitable.

5.2 If a test organism other than *Geobacillus stearothermophilus* or *Bacillus subtilis* ATCC 35021 (5230) is used, the suitability of the resistance of that test organism shall be determined.

NOTE For processes at less than 121 °C, microorganisms such as *Bacillus subtilis* ATCC 35021 (5230) could be used, particularly in sterilization of heat-sensitive liquids.

6 Suspension

The requirements of ISO 11138-1 apply.

7 Carrier and primary packaging

7.1 The suitability of the carrier and primary packaging materials for biological indicators for use in moist heat sterilization processes shall be demonstrated in accordance with the requirements of ISO 11138-1:2006, 5.2 and Annex B.

7.2 The exposure conditions for establishing compliance shall be:

- a) minimum exposure temperature: ≥ 5 °C above the manufacturer's stated maximum temperature;
- b) sterilizing agent: dry saturated steam; if the biological indicator is intended for use in a moist heat process not using dry saturated steam, e.g. an air/steam mixture, the appropriate air steam mixture should be used and noted as an exception to this part of ISO 11138;
- c) maximum exposure temperature: as stated by the manufacturer; if not stated by the manufacturer, a temperature of 140 °C shall be used;
- d) exposure time: ≥ 30 min.

NOTE These conditions have been selected to represent a realistic challenge to the carrier while remaining within the practical limits of a moist heat sterilization process.

8 Inoculated carriers and biological indicators

The requirements of ISO 11138-1 apply.

9 Population and resistance

9.1 The manufacturer shall state the resistance characteristics in accordance with ISO 11138-1:2006, 6.4.

9.2 The viable count shall be stated with increments $\leq 0,1 \times 10^n$ per unit, (e.g. per ml of suspension, per inoculated carrier or per biological indicator).

9.3 For inoculated carriers and biological indicators, the viable count shall be $\geq 1,0 \times 10^5$.

9.4 The resistance shall be expressed as the D value in minutes at 121 °C. The D value of each batch/lot of biological indicators or inoculated carriers shall be stated in minutes to one decimal place at 121 °C.

9.5 Suspensions, inoculated carriers or biological indicators containing *Geobacillus stearothermophilus* spores shall have a D_{121} value of $\geq 1,5$ min when tested according to the conditions given in Annex A. Other microorganisms shall have D values supporting the application. The z value of the test organisms in the suspension, on the inoculated carrier or in the biological indicator shall be determined at not less than three temperatures, in the range of 110 °C to 130 °C. These data shall be used to calculate the z value, which shall be ≥ 6 °C (see Annex B).

9.6 The resistance of a biological indicator may also be indicated by the term F_{BIO} value (see 11138-1:2006, 3.7).

The resistance characteristics specified in this part of ISO 11138 and any other part of ISO 11138 apply to the specific test conditions stated in the standards.

9.7 D values are determined according to methods given in Annexes C and D of ISO 11138-1:2006.

9.8 Determination of D value and survival-kill response characteristics requires the use of a resistometer applying the resistometer process parameters (see Annex A).

9.9 The survival-kill window can be calculated using the formulae in ISO 11138-1:2006, Annex E.

NOTE This information may be of value to the user when comparing different batches from the same manufacturer.

EXAMPLE

Using the formulae in ISO 11138-1:2006, Annex E with the minimum population and minimum D value requirements specified in this part of ISO 11138, the survival-kill response characteristics are:

— at 121 °C: survival time $\geq 4,5$ min and kill time $\leq 13,5$ min.

Annex A (normative)

Method for determination of resistance to moist heat sterilization

A.1 General

This method requires the use of a test apparatus referred to as a resistometer in this part of ISO 11138. The specifications of the resistometer process parameters for moist heat sterilization processes are provided in ISO 18472.

Specific requirements related to the test method are provided in A.2.

A.2 Method

A.2.1 Load the test samples on to suitable sample holders.

A.2.2 Pre-heat the resistometer chamber to the required operating temperature, e.g. $121\text{ }^{\circ}\text{C} \pm 0,5\text{ }^{\circ}\text{C}$.

A.2.3 Place the loaded sample holders in the chamber, close the chamber and initiate the process cycle.

A.2.4 Carry out the following sequence of operations:

- Step 1: Evacuate the chamber to a set point of $4,5\text{ kPa} \pm 1\text{ kPa}$ within 2 min.
- Step 2: Admit steam to the chamber to obtain the required temperature and pressure within 10 s. For the 0 minute exposure time, no steam shall be admitted.
- Step 3: Maintain these conditions for the required exposure time.
- Step 4: At the end of the exposure period, evacuate the chamber to 10 kPa or less within 1 min. The time taken to achieve a temperature of less than $100\text{ }^{\circ}\text{C}$ shall be less than 5 s. Then admit filtered air to ambient pressure.
- Step 5: At the end of the above process, remove the samples from the chamber, and cool down rapidly. Transfer the samples to the growth medium and incubate (see 11138-1:2006, Clause 7).

A.2.5 The transfer period should be documented and the same time period should be used for all tests.

A.3 Determination of resistance

Resistance characteristics shall be determined according to methods given in Annexes C, D and E of ISO 11138-1:2006.

Annex B (normative)

Calculation of z value and correlation coefficient, r^2

B.1 Using all the data obtained from either Annex C or D of ISO 11138-1:2006, plot the \log_{10} of the D value against exposure temperature in degrees Celsius. The z value is equal to the negative reciprocal of the slope of the best-fit rectilinear curve as determined by regression analysis.

NOTE See 9.5 for requirements regarding calculation of z value and correlation coefficient, r^2 .

B.2 The slope of the best-fit rectilinear curve is calculated using the following formula:

$$m = \frac{(nG) - (AB)}{(nC) - (A^2)}$$

where

m is the slope of the best-fit rectilinear curve;

n is the number of D value/temperature pairs;

$$G = \sum [t(\log_{10} y)];$$

$$A = \sum (t);$$

$$B = \sum (\log_{10} y);$$

$$C = \sum (t^2).$$

The data required for the calculation are given in Table B.1.

Table B.1 — Examples of data collected for regression analysis

D value (min) = y	Exposure temperature (°C) = t	$\log_{10} y$	t^2	$t(\log_{10} y)$	$(\log_{10} y)^2$
y_1	t_1	$\log_{10} y_1$	$(t_1)^2$	$t_1(\log_{10} y_1)$	$(\log_{10} y_1)^2$
y_2	t_2	$\log_{10} y_2$	$(t_2)^2$	$t_2(\log_{10} y_2)$	$(\log_{10} y_2)^2$
y_3	t_3	$\log_{10} y_3$	$(t_3)^2$	$t_3(\log_{10} y_3)$	$(\log_{10} y_3)^2$
y_n	t_n	$\log_{10} y_n$	$(t_n)^2$	$t_n(\log_{10} y_n)$	$(\log_{10} y_n)^2$
	$A = \sum_{i=1}^{i=n} t_i$	$B = \sum_{i=1}^{i=n} \log_{10} y_i$	$C = \sum_{i=1}^{i=n} (t_i)^2$	$G = \sum_{i=1}^{i=n} [t_i (\log_{10} y_i)]$	$E = \sum_{i=1}^{i=n} (\log_{10} y_i)^2$
Assigned variable	A	B	C	G	E

B.3 Table B.2 shows example calculations for the slope of the best-fit rectilinear curve.

Table B.2 — Examples of calculations for slope

<i>D</i> value (min) = <i>y</i>	Exposure temperature (°C) = <i>t</i>	$\log_{10} y$	t^2	$t(\log_{10} y)$	$(\log_{10} y)^2$
$y_1 = 2,0$	$t_1 = 121$	$\log_{10} y_1 = 0,301\ 0$	$(t_1)^2 = 146\ 41$	$t_1(\log_{10} y_1) = 36,421\ 0$	$(\log_{10} y_1)^2 = 0,090\ 6$
$y_2 = 1,1$	$t_2 = 124$	$\log_{10} y_2 = 0,041\ 4$	$(t_2)^2 = 153\ 76$	$t_2(\log_{10} y_2) = 5,133\ 6$	$(\log_{10} y_2)^2 = 0,001\ 7$
$y_3 = 0,4$	$t_3 = 129$	$\log_{10} y_3 = -0,397\ 9$	$(t_3)^2 = 166\ 41$	$t_3(\log_{10} y_3) = -51,329\ 1$	$(\log_{10} y_3)^2 = 0,158\ 3$
	$A = \sum_{i=1}^{i=3} t_i$	$B = \sum_{i=1}^{i=3} \log_{10} y_i$	$C = \sum_{i=1}^{i=3} (t_i)^2$	$G = \sum_{i=1}^{i=3} [t_i (\log_{10} y_i)]$	$E = \sum_{i=1}^{i=3} (\log_{10} y_i)^2$
Assigned variable	$A = 374$	$B = -0,055\ 5$	$C = 46\ 658$	$G = -9,774\ 5$	$E = 0,250\ 6$

$$m = \frac{(nG) - (AB)}{(nC) - (A^2)}$$

$$m = \frac{[(3)(-9,774\ 5)] - [(374)(-0,055\ 5)]}{[(3)(46\ 658)] - (374^2)}$$

$$m = \frac{(-29,323\ 5) - (-20,757\ 0)}{(139\ 974) - (139\ 876)}$$

$$m = \frac{-8,566\ 5}{98}$$

$$m = -0,087\ 4$$

B.4 The *z* value is equal to the negative reciprocal of the slope obtained and is calculated using the following formula:

$$z \text{ value} = -1 \left(\frac{1}{m} \right)$$

using the above calculated slope, the resulting *z* value is:

$$z = -1 \left(\frac{1}{-0,087\ 4} \right) = 11,441\ 6\ ^\circ\text{C rounded to one decimal point}$$

$$z = 11,4\ ^\circ\text{C}$$

B.5 The correlation coefficient, r^2 , for the linearity of the z value curve is calculated using the following formula:

$$r^2 = \frac{\{(G) - [(A)(B/n)]\}^2}{[(C) - (A^2/n)][(E) - (B^2/n)]}$$

where all variables are as defined in B.2 and $E = \sum (\log_{10} y)^2$.

B.6 Example calculations for the correlation coefficient for the linearity of the z value curve are given below.

Using the values from Table B.2,

$$r^2 = \frac{\{(-9,774.5) - [(374)(-0,0555/3)]\}^2}{[(46\,658) - (374^2/3)][(0,2506) - (-0,0555^2/3)]}$$

$$r^2 = \frac{[(-9,774.5) - [(-6,919.0)]]^2}{[(46\,658) - (46\,625,333/3)][(0,2506) - (0,0010)]}$$

$$r^2 = \frac{(-2,8555)^2}{(32,6667)(0,2496)}$$

$$r^2 = \frac{8,1539}{8,1536}$$

$$r^2 = 1,0000$$

Bibliography

- [1] ISO 14161, *Sterilization of health care products — Biological indicators — Guidance for the selection, use and interpretation of results*
- [2] ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

Preview Only

Acknowledgements

Members of Technical Committee on Quality Management and Corresponding General Aspects for Medical Devices

Dr Mohd Shah Dato' Idris (Chairman)	Ministry of Health Malaysia
Ms Salbiah Yaakop (Secretary)	SIRIM Berhad
Dr T Mahadevan	Association of Private Hospitals of Malaysia
Haji Tauran Zaidi Ahmad Zaidi/	Biomedical Engineering Association Malaysia
Mr Khairul Azmy Kamaluddin	
Mr Ho Yeam Chan	Federation of Malaysian Manufacturers
Ms Yap Siow Ling	Institute of Quality Malaysia
Mr Law Soo Chin	Malaysia Medical Device Association
Dr Kerpai Singh Pannu	Malaysian Medical Association
Dr Wan Manshol Wan Zin/	Malaysian Nuclear Agency
Ms Rohaizah Ahmad	
Dr Abdul Kadir Mohamed	Malaysian Rubber Export Promotion Council
Dr Tan Ah Seng	Malaysian Rubber Glove Manufacturers' Association
	SIRIM QAS International Sdn Bhd
Mr Chang Boon Ping	SterilGamma (M) Sdn Bhd
Ms Norhayati Hussin	

Preview Only

© Copyright 2010

All rights reserved. No part of this publication may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the Department of Standards Malaysia.