



MALAYSIAN STANDARD

MS CISPR 16-1-1:2012

**Specification for radio disturbance and immunity measuring apparatus and methods
- Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus
(Second revision)
(CISPR 16-1-1:2010, IDT)**

ICS: 33.100.10

Descriptors: specification, radio disturbance, immunity, measuring apparatus, method

© Copyright 2012

DEPARTMENT OF STANDARDS MALAYSIA

DEVELOPMENT OF MALAYSIAN STANDARDS

The **Department of Standards Malaysia (STANDARDS MALAYSIA)** is the national standards and accreditation body of Malaysia.

The main function of STANDARDS MALAYSIA is to foster and promote standards, standardisation and accreditation as a means of advancing the national economy, promoting industrial efficiency and development, benefiting the health and safety of the public, protecting the consumers, facilitating domestic and international trade and furthering international cooperation in relation to standards and standardisation.

Malaysian Standards (MS) are developed through consensus by committees which comprise balanced representation of producers, users, consumers and others with relevant interests, as may be appropriate to the subject at hand. To the greatest extent possible, Malaysian Standards are aligned to or are adoption of international standards. Approval of a standard as a Malaysian Standard is governed by the Standards of Malaysia Act 1996 [Act 549]. Malaysian Standards are reviewed periodically. The use of Malaysian Standards is voluntary except in so far as they are made mandatory by regulatory authorities by means of regulations, local by-laws or any other similar ways.

STANDARDS MALAYSIA has appointed **SIRIM Berhad** as the agent to develop, distribute and sell the Malaysian Standards.

For further information on Malaysian Standards, please contact:

Department of Standards Malaysia
Ministry of Science, Technology and Innovation
Level 1 & 2, Block 2300, Century Square
Jalan Usahawan
63000 Cyberjaya
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 8318 0002
Fax: 60 3 8319 3131
<http://www.standardsmalaysia.gov.my>

E-mail: central@standardsmalaysia.gov.my

OR **SIRIM Berhad**
(Company No. 367474 - V)
1, Persiaran Dato' Menteri
Section 2, P.O. Box 7035
40700 Shah Alam
Selangor Darul Ehsan
MALAYSIA

Tel: 60 3 5544 6000
Fax: 60 3 5510 8095
<http://www.sirim.my>

E-mail: msonline@sirim.my

CONTENTS

Committee representation.....v

National foreword.....vi

Foreword.....vii

Introduction.....ix

1 Scope.....9

2 Normative references9

3 Terms and definitions10

4 Quasi-peak measuring receivers for the frequency range 9 kHz to 1 000 MHz.....13

 4.1 General.....13

 4.2 Input impedance.....13

 4.3 Sine-wave voltage accuracy.....13

 4.4 Response to pulses.....13

 4.4.1 Amplitude relationship (absolute calibration).....13

 4.4.2 Variation with repetition frequency (relative calibration).....14

 4.5 Selectivity18

 4.5.1 Overall selectivity (passband).....18

 4.5.2 Intermediate frequency rejection ratio.....19

 4.5.3 Image frequency rejection ratio.....19

 4.5.4 Other spurious responses.....20

 4.6 Limitation of intermodulation effects21

 4.7 Limitation of receiver noise and internally generated spurious signals22

 4.7.1 Random noise22

 4.7.2 Continuous wave22

 4.8 Screening effectiveness22

 4.8.1 General22

 4.8.2 Limitation of radio-frequency emissions from the measuring receiver.....23

 4.9 Facilities for connection to a discontinuous disturbance analyzer23

5 Measuring receivers with peak detector for the frequency range 9 kHz to 18 GHz23

 5.1 General.....23

 5.2 Input impedance.....23

 5.3 Fundamental characteristics.....24

 5.3.1 Bandwidth24

 5.3.2 Charge and discharge time constants ratio24

 5.3.3 Overload factor.....25

 5.4 Sine-wave voltage accuracy.....25

 5.5 Response to pulses.....25

 5.6 Selectivity25

 5.7 Intermodulation effects, receiver noise, and screening26

6 Measuring receivers with average detector for the frequency range 9 kHz to 18 GHz.....27

 6.1 General.....27

 6.2 Input impedance.....27

 6.3 Fundamental characteristics.....27

 6.3.1 Bandwidth27

 6.3.2 Overload factor.....27

 6.4 Sine-wave voltage accuracy.....28

 6.5 Response to pulses.....28

 6.5.1 General.....28

6.5.2	Amplitude relationship	28
6.5.3	Variation with repetition frequency.....	29
6.5.4	Response to intermittent, unsteady and drifting narrowband disturbances.....	29
6.6	Selectivity	31
6.7	Intermodulation effects, receiver noise, and screening	31
7	Measuring receivers with rms-average detector for the frequency range 9 kHz to 18 GHz.....	31
7.1	General.....	31
7.2	Input impedance.....	31
7.3	Fundamental characteristics.....	32
7.3.1	Bandwidth	32
7.3.2	Overload factor.....	32
7.4	Sine-wave voltage accuracy	33
7.5	Response to pulses.....	33
7.5.1	Construction details.....	33
7.5.2	Amplitude relationship	33
7.5.3	Variation with repetition frequency.....	34
7.5.4	Response to intermittent, unsteady and drifting narrowband disturbances.....	34
7.6	Selectivity	35
7.7	Intermodulation effects, receiver noise, and screening	35
8	Measuring receivers for the frequency range 1 GHz to 18 GHz with amplitude probability distribution (APD) measuring function.....	35
9	Disturbance analyzers	36
9.1	General.....	36
9.2	Fundamental characteristics.....	37
9.3	Test method for the validation of the performance check for the click analyzer	43
9.3.1	Basic requirements.....	43
9.3.2	Additional requirements.....	44
Annex A (normative)	Determination of response to repeated pulses of quasi-peak and rms-average measuring receivers (See 3.6, 4.4.2, 7.3.2 and 7.5.1)	45
Annex B (normative)	Determination of pulse generator spectrum (See 4.4, 5.5, 6.5, 7.5).....	51
Annex C (normative)	Accurate measurements of the output of nanosecond pulse generators (See 4.4, 5.5, 6.5, 7.5)	53
Annex D (normative)	Influence of the quasi-peak measuring receiver characteristics on its pulse response (See 4.4.2).....	55
Annex E (normative)	Response of average and peak measuring receivers (See 6.3.1)	56
Annex F (normative)	Performance check of the exceptions from the definitions of a click according to 4.2.3 of CISPR 14-1	65
Annex G (informative)	Rationale for the specifications of the APD measuring function	72
Annex H (informative)	Characteristics of a quasi-peak measuring receiver.....	75
Annex I (informative)	Example of EMI receiver and swept spectrum analyzer architecture.....	76
	Bibliography.....	78
	Figure 1 – Pulse response curves	17

Figure 2 – Limits of overall selectivity	20
Figure 3 – Arrangement for testing intermodulation effects	21
Figure 4 – Limits for the overall selectivity – pass band (Band E).....	26
Figure 5 – Block diagram of an average detector.	30
Figure 6 – Screenshot showing response of the meter-simulating network to an intermittent narrowband signal.....	30
Figure 7 – Example of a disturbance analyzer.....	38
Figure 8 – A graphical presentation of test signals used in the test of the analyzer for the performance check against the definition of a click according to Table 14	39
Figure E.1 – Correction factor for estimating the ratio B_{imp}/B_6 for other tuned circuits	57
Figure E.2 – Pulse rectification coefficient P	59
Figure E.3 – Example (spectrum screenshot) of a pulse-modulated signal with a pulse width of 200 ns	60
Figure E.4 – Pulse-modulated RF signal applied to a measuring receiver	61
Figure E.5 – Filtering with a B_{imp} much smaller than the prf	61
Figure E.6 – Filtering with a B_{imp} much wider than the prf.....	62
Figure E.7 – Calculation of the impulse bandwidth	62
Figure E.8 – Example of a normalized linear selectivity function	64
Figure F.1 – A graphical presentation of the test signals used for the performance check of the analyzer with the additional requirements according to Table F.1.....	71
Figure G.1 – Block diagram of APD measurement circuit without A/D converter.....	73
Figure G.2 – Block diagram of APD measurement circuit with A/D converter.....	73
Figure G.3 – Example of display of APD measurement	74
Figure I.1 – Example block diagram of EMI receiver consisting of swept spectrum analyzer with added preselector, preamplifier and quasi-peak/average detector	76
Table 1 – Test pulse characteristics for quasi-peak measuring receivers (see 4.4.1)	13
Table 2 – Pulse response of quasi-peak measuring receivers	18
Table 3 – Combined selectivity of CISPR measuring receiver and high-pass filter.....	19
Table 4 – Bandwidth characteristics for intermodulation test of quasi-peak measuring receivers (see 4.6).....	22
Table 5 – VSWR requirements for receiver input impedance.....	24
Table 6 – Bandwidth requirements for measuring receivers with peak detector	24
Table 7 – Relative pulse response of peak and quasi-peak measuring receivers for the same bandwidth (frequency range 9 kHz to 1 000 MHz).....	25
Table 8 – Bandwidth requirements for measuring receivers with average detector	27
Table 9 – Relative pulse response of average and quasi-peak measuring receivers for the same bandwidth (frequency range 9 kHz to 1 GHz).....	28
Table 10 – Maximum reading of average measuring receivers for a pulse-modulated sine-wave input in comparison with the response to a continuous sine wave having the same amplitude	30
Table 11 – VSWR requirements of input impedance.....	32
Table 12 – Bandwidth requirements for measuring receivers with rms-average detector	32
Table 13 – Minimum pulse repetition rate without overload	32
Table 14 – Relative pulse response of rms-average and quasi-peak measuring receivers.....	33

Table 15 – Pulse response of rms-average measuring receiver	34
Table 16 – Maximum reading of rms-average measuring receivers for a pulse-modulated sine-wave input in comparison with the response to a continuous sine wave having the same amplitude	35
Table 17 – Disturbance analyzer performance test – Test signals used for the check against the definition of a click.....	40
Table B.1 – Pulse generator characteristics	51
Table E.1 – B_{imp} and A_{imp} values for a peak measuring receiver	58
Table E.2 – Carrier level for pulse-modulated signal of 1,4 nVs	60
Table F.1 – Disturbance analyzer test signals ^a	66
Table H.1 – Characteristics of quasi-peak measuring receivers	75

Committee representation

The Industry Standards Committee on Generation, Transmission and Distribution of Energy (ISC E) under whose authority this Malaysian Standard was adopted, comprises representatives from the following organisations:

Association of Consulting Engineers Malaysia
Atomic Energy Licensing Board
Department of Standards Malaysia
Federation of Malaysian Manufacturers
Jabatan Kerja Raya Malaysia
Malaysian Association of Standards Users
Malaysian Cable Manufacturers Association
Malaysian Electrical Appliances and Distributors Association
Malaysian Green Technology Corporation
Ministry of Domestic Trade, Co-operatives and Consumerism
Ministry of International Trade and Industry
Persatuan Kontraktor Elektrikal dan Mekanikal Melayu Malaysia
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Komunikasi dan Multimedia Malaysia
Suruhanjaya Tenaga
Tenaga Nasional Berhad
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
Universiti Teknologi Malaysia

The Technical Committee on Electromagnetic Compatibility which supervised the adoption of the IEC Standard as Malaysian Standard consists of representatives from the following organisations:

Federation of Malaysian Manufacturers
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Komunikasi dan Multimedia Malaysia
Suruhanjaya Tenaga
Tenaga Nasional Berhad
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
TNB Research Sdn Bhd
Universiti Teknologi Malaysia

The Working Group on EMC-CISPR-IEC which recommended the adoption of the IEC Standard as Malaysian Standard consists of representatives from the following organisations:

Institut Penyelidikan Sains dan Teknologi Pertahanan
SIRIM Berhad (Secretariat)
SIRIM QAS International Sdn Bhd
Suruhanjaya Tenaga
The Electrical and Electronics Association of Malaysia
The Institution of Engineers, Malaysia
TNB Research Sdn Bhd

NATIONAL FOREWORD

The adoption of the IEC Standard as a Malaysian Standard was recommended by the Working Group on EMC-CISPR-IEC under the authority of the Industry Standards Committee on Generation, Transmission and Distribution of Energy.

This Malaysian Standard is the second revision of MS CISPR 16-1-1, *Specification for radio disturbance and immunity measuring apparatus and methods, Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus*.

This Malaysian Standard is identical with CISPR 16-1-1:2010, *Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus*, published by the International Electrotechnical Commission (IEC). However, for the purposes of this Malaysian Standard, the following apply:

- a) in the source text, "this International Standard" should read "this Malaysian Standard";
and
- b) the comma which is used as a decimal sign (if any), to read as a point.

This standard cancels and replaces MS CISPR 16-1-1:2010.

Compliance with a Malaysian Standard does not of itself confer immunity from legal obligations.

NOTE. IDT on the front cover indicates an identical standard i.e. a standard where the technical content, structure, and wording (or is an identical translation) of a Malaysian Standard is exactly the same as in an International Standard or is identical in technical content and structure although it may contain the minimal editorial changes specified in clause 4.2 of ISO/IEC Guide 21-1.